



Bio-Rad Laboratories, Inc.
Certificate of Conformance

Material Description: Gel Filtration Standard, pkg 6
Material Number: 1511901
Batch Number: 64022360
Manufacture Date: 2015-09-09
Expiration Date: 2018-08-06

This material meets all established quality control specifications.

Components included in Kit:

64016039 GEL FILTRATION STD.,LABELED,1

Shamla S. Bradford
Prepared by

2015-10-07
Date



Bio-Rad Laboratories, Inc.
Certificate of Analysis

Material Description: GEL FILTRATION STD.,LABELED,1
Material Number: 9700566
Batch Number: 64016039
Manufacture Date: 2015-08-07
Expiration Date: 2018-08-10

<u>Characteristic</u>	<u>Specification</u>		<u>Results</u>	
	Lower Limit	Upper Limit		
Peak Ht of Thyroglobulin as % of Vitamin B12	30		83 %	Pass
Peak Ht of IgG as % of Vitamin B12	30		90 %	Pass
Peak Ht of Ovalbumin as % of Vitamin B12	30		68 %	Pass
Peak Ht of Myoglobin as % of Vitamin B12	70		109 %	Pass
Thyroglobulin & IgG Origin: must be USA			Pass	Pass

Shamla S. Bradford
Prepared by

2015-10-07
Date